

# Mouse IgG-heavy and light chain cross-adsorbed Antibody

Donkey Polyclonal

Conjugate Alkaline Phosphatase

Antigen Affinity Purified

Catalog No. A90-337AP

Lot No. A90-337AP-4



<b>APPLICATIONS</b>	WB, IHC, ICC, ELISA
<b>SPECIES REACTIVITY</b>	Mouse. Minimum reactivity to bovine, chicken, goat, human, rabbit, rat and sheep
<b>ISOTYPE</b>	IgG
<b>AMOUNT</b>	1 ml at 0.5 mg/ml
<b>STORAGE/SHELF LIFE</b>	2 - 8° C / 1 year from date of receipt
<b>PHYSICAL STATE</b>	Liquid
<b>BUFFER</b>	50 mM HEPES pH 7.1, 0.1 M NaCl, 1 mM MgCl <sub>2</sub> , 0.1 mM ZnCl <sub>2</sub> containing 0.2% BSA and 0.09% NaN <sub>3</sub>
<b>ORIGIN</b>	USA
<b>PRODUCTION PROCEDURES</b>	Antiserum was cross adsorbed using bovine, chicken, goat, human, rabbit, rat and sheep immunosorbents to remove cross reactive antibodies. The antibody to mouse IgG was isolated by affinity chromatography using antigen coupled to agarose beads and conjugated to alkaline phosphatase.

Antibody concentration was determined by extinction coefficient: absorbance at 280 nm of 1.4 equals 1.0 mg of IgG.

By immunoelectrophoresis and ELISA this antibody reacts specifically with mouse IgG and with light chains common to other mouse immunoglobulins. No antibody was detected against non-immunoglobulin serum proteins. Less than 2% cross reactivity to bovine, chicken, goat, human, rabbit, rat and sheep IgG was detected. This antibody may cross react with IgG from other species.

**APPLICATIONS** Centrifuge tube to remove product from lid. Optimal working dilutions should be determined experimentally by the investigator. Prepare working dilution immediately before use.

Western Blot	1:2,500 - 1:25,000
Immunohistochemistry	1:100 - 1:1,000
Immunocytochemistry	1:100 - 1:1,000
ELISA	1:2,500 - 1:25,000

**APPLICATION NOTES** Not all listed applications have been specifically tested by our laboratory.

**ADDITIONAL INFO** <https://www.bethyl.com/product/A90-337AP>  
Use the link above to view SDS, a current list of citations, and other product specific information.

This document certifies that this product has met all of the quality control standards defined by Bethyl Laboratories, Inc.

Eric McIntush, PhD | Chief Scientific Officer

Date: January 28, 2019