



Diacclone

DIACLONE Immunology Products

CD106 ELISA Kit

Human CD106 ELISA kit

CATALOGUE N° :	850.580.048, 850.580.096, 850.580.192
SPECIFICITY :	Recognizes both natural and recombinant human soluble VCAM-1
RANGE :	1.56 ng / ml - 50 ng / ml
SENSITIVITY :	<0.6 ng / ml
INCUBATION :	1h40
SAMPLE TYPES :	Serum Plasma Cell culture supernatant
SAMPLE SIZE :	100µl
CROSS REACTION :	No cross reactivity with other human adhesion molecules
KIT SIZE :	48 (Discovery), 96, 192 tests
REFERENCES :	Atteritano, M. et al., J Clin Endocrinol Metab.,2007;92: 3068 - 3075 Castilla, P. et al., Am J Clin Nutr.,(2006; 84(1): 252-62. Castilla, P. et al., Am. J. Clinical Nutrition,2008; 87(4): 1053-1061. Chen M-C. et al., Chest, 2004; 125(4): 1213 - 1217 Connolly, G. M. et al., BMC Nephrol.,2011; 12: 23. Cosentino, F. et al., Stroke,2009; 40(1): 306-308. Diamanti-Kandarakis, E. et al., Hum Reprod., 2006; 21(6): 1426-31. Ehlermann, P. et al.,Cardiovasc Diabetol.,2006; 5: 6 Engelen, L., et al., Eur. J. Endocrinol.,2011; 164(3): 371-379 Foussat A. et al., Blood, 2001; 98(6): 1678 - 1686 Galea, P. et al., Res Immunol., 1997; 148(2): 109-17. Grooteman, M. P. et al., Nephrol Dial Transplant, 2005; 20(12): 2751-8. Hartweg, J. et al., Clin. Chem.,2007; 53(10): 1858-1860. Hetteema, M. E. et al., Arthritis Res Ther.,2008; 10(2): R49 Hidalgo, M. et al.,Genes Nutr.,2012; 7(2): 295-306 Ho, C. Y. et al., Rheumatology(Oxford),2003; 42(1): 117-22. Janowska, J. et al., Int J Med Sci.,2016;13(7): 490-9. Jha, H. C. et al., Mediators Inflamm., 2009: 561532 Jie, K. E. et al.,Nephrol. Dial. Transplant.,2010; 25(6): 1875-1882. Just, N. et al., Clin Exp Immunol.,2003; 132(1): 61-9. Larsson, S. et al., Int J Chron Obstruct Pulmon Dis.,2011;6: 83-8 le Poole, C. Y. et al.,Perit. Dial. Int.,2013;33(2): 208-212 Liao, B. et al., Emerg Microbes Infect.,2015;4(4):e24 Lieuw-A-Fa,ML. et al., Nephrol. Dial. Transplant., 2004; 19(3): 631 - 636 Llorente, L. et al., Arthritis Rheum.,2000; 43(8): 1790-800. Lund, S.S. Et al., Eur. J. Endocrinol.,2008;158(5): 631-641. Manneras-Holm, L. et al., J. Clin. Endocrinol. Metab.,2011; 96(4): 1068-1076. Monchanin, G. et al., J Appl Physiol., 2007; 102(1): 169-73. Nebor, D. et al., Haematologica,2011;96(11):1589-1594 Pattan, V. et al., J Clin Med Res.,2015; 7(8): 613-9 Persson, F. et al., Diabetes,2006; 55(12): 3550-5. Persson, F. et al.,Diabetes Care,2009; 32(10): 1873-1879. Sammons J. et al., Stem Cells 2000; 18(3): 214 - 219 Schalkwijk C.G. et al., Ann. N.Y. Acad. Sci, 2005; 1043: 662 - 670 Singh, A. et al., Diabetes Care, 2012; 35: 1322 - 1327 Singhania, N. et al., QJM,2008; 101(6): 449-455. Stam, F. et al., Nephrol Dial Transplant,2003; 18(5): 892-8. Tanwar, R. S. et al., Evid Based Complement Alternat Med; 2011: 127641 Van Dam, B. et al.,Cardiovasc Res.,2003; 57(2): 563-571. Van de Kerkhof J. et al., Nephrol. Dial. Transplant., 2004; 19(4): 910 - 916 van Ree, R. M. et al., Nephrol Dial Transplant,2005; 20(11): 2524-31. Wijnstok, N. J. et al., J Adolesc Health,2010; 47(4): 346-51. Yao, G. H. et al., Lupus,2008;17(8): 720-726. Zedan, M. et al., ISRN Pediatr., 2013: 824781

KIT CONTENT : Pre-coated 12 strip plates, biotinylated secondary antibody, standards, controls (when available), buffers, Streptavidin-HRP, TMB, Stop Reagent

Version 17 - 06.17

For research use only

25 YEARS OF EXPERTISE

With over 25 years experience and extensive expertise we are committed to providing excellence in Monoclonal Antibody and Immunoassay development.

The expanding range of Diacclone Immunology products is specifically designed to advance research applications.

Our experience and expertise coupled to the diversity of our product range makes Diacclone a clear choice to

Fast Track Your Research...

DIACLONE PRODUCTS



Monoclonal Antibodies



ELISA



ELISpot



DIAPlex

DIACLONE SAS

6 rue Dr Jean-François-Xavier Girod
- BP 1985
25020 Besançon cedex - France
Tel: +33 (0)3 81 41 38 38
Fax: +33 (0)3 81 41 36 36
info@diacclone.com



www.diacclone.com