



Diacclone

DIACLONE Immunology Products

IL-8 ELISA Kit Human IL-8 ELISAKit

CATALOGUE N° :	950.050.048, 950.050.096, 950.050.192
SPECIFICITY :	Recognizes both natural and recombinant human IL-8
RANGE :	62.5 pg / ml - 2000 pg / ml
SENSITIVITY :	< 25 pg / ml
INCUBATION :	1h40
SAMPLE TYPES :	Serum Plasma Cell culture supernatant
SAMPLE SIZE :	100 µl
CROSS REACTION :	No cross reactivity with other human cytokines
KIT SIZE :	48 (Discovery), 96, 192 tests

- Akbari, F. et al., Iran J Neurol.,2016; 15(2): 75-9.
Audibert, C. et al., Infect Immun., 2001; 69(3): 1625-9.
Bäckhed F. et al., J. Biol. Chem., 2002; 277(20): 18198 - 19205
Backhed, F. et al., Infect Immun., 2003; 71(6): 3357-60.
Bellanger, A-P. et al., Clin. Vaccine Immunol., 2013 ; 20(8): 1133-1142.
Bonnans, C. et al., Am J Pathol.,2006;168(4):1064-72.
Bote, M. E. et al., PLoS One,2013 ; 8(9): e74524
Chabbert-de Ponnat, I. et al., Int Immunol.,2005; 17(4): 439-47.
Chavushyan, A. et al., Schizophr Res Treatment,2013: 125264
Coconnier, M. H. et al., Appl Environ Microbiol.,2000; 66(3): 1152-7.
Eilertsen, G. O. et al., Lupus,2011 ; 20(6): 607-613.
Eleftheriadis, T. et al., Hippokratia,2011;15(3): 238-43
Elmerich S. et al., Carcinogenesis, 2000; 21(4): 753 -756
Fan, H. et al., Mbl Immunol.,2007; 44(12): 3092-9
Gehrke, T. et al., Oncol Lett.,2016; 12(5): 3549-3554
Giamarellos-Bourboulis, E. J. et al., Crit Care,2006; 10(3): R76
Gronella J. et al., Gut, 2005; 54(9): 1244 - 1253
Gombocz, K. et al., Crit Care,2007; 11(4): R87
Guerrero, A. et al., Aging Cell,2015 ; 14(2): 274-83
Hot, A. et al., Ann Rheum Dis, 2012; 71: 768 - 776.
Krishna, S. et al., PLoS One,2011; 6(12): e28476
Krzysiek, R. et al., J Immunol.,1999;162(8): 4455-63
Kumpf, O. et al., Crit Care,2010; 14(3): R103.
Lee, H. W. et al., Ann Dermatol.,2013 ; 25(2): 173-80
Lekkou A. et al., Clin. Diagn. Lab. Immunol., 2004; 11(1):161 - 167
Levy, B. D. et al., Prostaglandins Leukot Essent Fatty Acids,2011; 84(1-2): 43-50.
Maruyama, H. et al., Lipids Health Dis.,2014; 13: 78
Mazurkiewicz P. et al., Mbl Microbiol.,2008; 67(6): 1371-83.
Rajappa, M. et al., Angiology,2009; 60(4): 419-426.
Rana, S. V. et al., J Crohns Colitis,2014 ; 8(8): 859-865
Routsis, C. et al., Clin Exp Immunol.,2005;142(1): 62-7.
Roy, S. et al., Immunology,2004; 112(3): 471-80.
Rybka, J. et al., Clin Exp Med.,2015; 16(4): 493-502.
Saraiva M. et al., JEM2002; 196(6): 829-839
Sell, H. et al., Am J Physiol Endocrinol Metab., 2008;294:E1070-E1077
Sjolinder, H. et al., Infect Immun., 2008; 76(9):3959-66
Tenca C. et al., 2005; 174(11): 6757 - 6763
Toufiraï, O. et al., Clin Exp Immunol.,2007; 149(2): 372-7
Tsaganos, T. et al., BMC Infect Dis.,2006; 6: 142.
Tutar, E. et al., World J Gastroenterol.,2008;14(20): 3218-23.
van de Sande, W.W.J. et al., J. Immunol.,2007; 179(5): 3065-3074.
Vassallo, R. et al., Mbl Immunol.,2008; 45(12): 3321-9.

Yadav, U. C. S. et al., J. Immunol.,2009; 183(7): 4723-4732.
Yadav, U. C. et al., PLoS One,2009; 4(8): e6535
Zrioual, S. et al., J. Immunol.,2008; 180(1): 655-663.

KIT CONTENT : Pre-coated 12 strip plates, biotinylated secondary antibody, standards, controls (when available), buffers, Streptavidin-HRP, TMB, Stop Reagent.

25 YEARS OF EXPERTISE

With over 25 years experience and extensive expertise we are committed to providing excellence in Monoclonal Antibody and Immunoassay development.

The expanding range of Diacclone Immunology products is specifically designed to advance research applications.

Our experience and expertise coupled to the diversity of our product range makes Diacclone a clear choice to

Fast Track Your Research...

DIACLONE PRODUCTS



Monoclonal Antibodies



ELISA



ELISpot



DIAPlex

DIACLONE SAS

6 rue Dr Jean-François-Xavier Girod
- BP 1985
25020 Besançon cedex - France
Tel: +33 (0)3 81 41 38 38
Fax: +33 (0)3 81 41 36 36
info@diacclone.com



www.diacclone.com

